

Participant No.

2010 AUTO DIFFERENTIAL MODULE 2

CELL DYN 3000, 3200/Ruby, 3500, 3700, 4000/Sapphire

PLEASE SEE REVERSE FOR INSTRUCTIONS

INSTRUMENT(Manufacturer/Model): _____		SERIAL NUMBER: _____	
DUE DATE	SAMPLE NUMBER		UNITS
□□/□□	AD10-2-□□a	AD10-2-□□b	
WCC	□□□.□	□□□.□	x 10 ⁹ /L
NEUT / GRAN	□□.□	□□.□	%
LYMP	□□.□	□□.□	%
MONO	□□.□	□□.□	%
EOS	□□.□	□□.□	%
BASO	□□.□	□□.□	%

DATE RECEIVED: □□ □□ □□ DATE RETURNED: □□ □□ □□

AUTOMATED DIFFERENTIAL (AD2) PROCESSING INSTRUCTIONS

These directions have been obtained from the manufacturer's insert for the control material supplied.

- Remove the vials of control from the refrigerator and warm to room temperature (18°C to 30°C) for 15 minutes before use.
 - To mix (Do not mix mechanically).
 - Hold the vial horizontally between the palms of the hands and roll the vial back and forth for 20 to 30 seconds. Do not shake.
 - Mix by rapid inversion until all cells are resuspended.
 - Vials stored for an extended period of time may need extra mixing.
 - Gently invert the vial 8 to 10 times immediately before sampling.
 - Refer to the instrument manual for the system in use for analysing control material.
 - After sampling, return to refrigeration for maximum open-vial stability. If run in the open mode, wipe the threads of both the vial and cap before replacing the cap and returning to refrigeration.
-